

INTENDED USE

The SPIFE Nexus ImmunoFix method is a fully automated solution intended for the qualitative identification of monoclonal gammopathies in serum and urine using protein electrophoresis and immunofixation on the SPIFE Nexus system.

SUMMARY

Immunofixation electrophoresis (IFE) is a two-stage procedure using agarose gel high resolution electrophoresis in the first stage and immunoprecipitation in the second. There are numerous applications for IFE in research, forensic medicine, genetic studies and clinical laboratory procedures. The greatest demand for IFE is in the clinical laboratory where it is primarily used for the detection of monoclonal gammopathies. A monoclonal gammopathy is a primary disease state in which a single clone of plasma cells produces elevated levels of an immunoglobulin of a single class and type. Such immunoglobulins are referred to as monoclonal proteins, M-proteins or paraproteins. Their presence may be of a benign nature or of uncertain significance. In some cases, they are indicative of a malignancy such as multiple myeloma or Waldenstrom's macroglobulinemia. Differentiation must be made between polyclonal and monoclonal gammopathies because polyclonal gammopathies are only a secondary disease state due to clinical disorders such as chronic liver diseases, collagen disorders, rheumatoid arthritis and chronic infections.

IFE has been used for the study of immunoglobulins since 1976.^{1,2} The SPIFE Nexus IFE method offers hands-free operation, ease of interpretation, excellent resolution, reagent conservation and rapid turnaround.

PRINCIPLE

Proteins are first resolved by electrophoresis. In the second stage, the soluble antigen and antibody are allowed to react. The resultant antigen-antibody complex(es) may become insoluble (as long as the antibody is in slight excess or near equivalency) and precipitate. The precipitation rate depends on the proportions of the reactants, temperature, salt concentration and the pH of the solution. The unreacted proteins are removed by a washing step and the antigen-antibody complex (which might be visible as a white cloudy band in the unstained gel against a dark background) is visualized by staining. The bands in the protein separation are compared with the precipitin bands obtained with immunofixation.

REAGENT

1. SPIFE IFE-6 Gel

Ingredients: Each gel contains agarose in tris-barbital/MOPS buffer with a stabilizer and a preservative.

WARNING: FOR IN-VITRO DIAGNOSTIC USE ONLY. CAUTION: DO NOT INGEST. The gel contains barbital which, in sufficient quantity, can be toxic.

Preparation for Use: The gels are ready for use as packaged.

Storage and Stability: The gels should be stored horizontally at room temperature (15 to 30°C) and are stable until the expiration date indicated on the package. The gels must be stored in the protective packaging in which they are shipped. **DO NOT REFRIGERATE OR FREEZE.**

Signs of Deterioration: Any of the following conditions may indicate deterioration of the gel: (1) crystalline appearance indicating the agarose has been frozen, (2) cracking and peeling indicating drying of the agarose, (3) bacterial growth indicating contamination, (4) thinning of gel blocks.

2. SPIFE Nexus Violet

Ingredients: The stain is comprised of 0.2% (w/v) acid violet stain and 10% acetic acid.

WARNING: FOR IN-VITRO DIAGNOSTIC USE ONLY. DO NOT INGEST.

Preparation for Use: The stain is ready for use as packaged.

Storage and Stability: The stain solution is stable for one year when stored at 15 to 30°C in a closed container.

Signs of Deterioration: The stain should be a homogeneous mixture free of precipitate.

3. Citric Acid Destain

Ingredients: After dissolution, the destain contains 0.3% (w/v) citric acid.

WARNING: FOR IN-VITRO DIAGNOSTIC USE. DO NOT INGEST—IRRITANT.

Preparation for Use: Pour 11 L of deionized water into the Destain vat. Add the entire package of Destain and mix until completely dissolved.

Storage and Stability: Store the Destain at 15 to 30°C. It is stable until the expiration date on the package.

Signs of Deterioration: Discard if solution becomes cloudy.

4. Clear Wash

Ingredients: The powder contains anionic and nonionic surfactants, sodium carbonate, enzymes and sodium chloride.

WARNING: FOR IN-VITRO DIAGNOSTIC USE

Preparation for Use: Dissolve powder in 8 L of deionized water and mix thoroughly.

Storage and Stability: Store the dry powder at 15 to 30°C until the expiration date indicated on the label. The buffer solution should be stored at 15 to 30°C.

Signs of Deterioration: Discard if signs of bacterial contamination are seen.

5. SPIFE Nexus Pipette Wash

Ingredients: The buffer solution contains a sodium hydroxide solution.

WARNING: FOR IN-VITRO DIAGNOSTIC USE. DANGER: CORROSIVE—NEVER PIPETTE BY MOUTH. DO NOT INGEST.

Preparation for Use: The buffer solution is ready for use as packaged.

Storage and Stability: The buffer solution should be stored at 15 to 30°C and is stable until the expiration date indicated on the vial.

Signs of Deterioration: The buffer solution should be a clear solution.

6. SPIFE Nexus IFE Protein Fixative

Ingredients: The fixative contains 4.0% sulfosalicylic acid, 6.7% trichloroacetic acid, 0.2% glutaraldehyde and 1.7% guanidine HCl.

WARNING: FOR IN-VITRO DIAGNOSTIC USE. CORROSIVE—NEVER PIPETTE BY MOUTH. DO NOT INGEST.

Preparation for Use: The fixative is ready for use as packaged.

Storage and Stability: The fixative should be stored at 2 to 8°C and is stable until the expiration date indicated on the vial.

Signs of Deterioration: The fixative should be a clear solution.

7. Antisera to Human IgG, IgA, IgM, Kappa Light Chain and Lambda Light Chain

Ingredients: Antisera vials in the kit contain monospecific antisera to human immunoglobulin heavy chains, IgG, IgM, IgA and to human light chains, Kappa and Lambda. The antisera have been prepared in goat. Each vial of antiserum contains a stabilizer and sodium azide as a preservative.

WARNING: FOR IN-VITRO DIAGNOSTIC USE ONLY. To prevent the formation of toxic vapors, do not mix with acidic solutions. When discarding, always flush sink with copious amounts of water. This will prevent the formation of metallic azides which, when highly concentrated in metal plumbing, are potentially explosive. In addition to purging pipes with water, plumbing should occasionally be decontaminated with 10% NaOH.

Preparation for Use: The antisera are ready for use as packaged.

Storage and Stability: The antisera should be stored at 2 to 8°C and are stable until the expiration date indicated on the vial.

Signs of Deterioration: Extremely cloudy antisera may be indicative of bacterial contamination.

INSTRUMENT

A SPIFE Nexus analyzer must be used to apply samples, electrophorese, apply antisera and fixative, wash, stain, destain, dry and then scan the gel. Gels may also be scanned on a separate densitometer such as QuickScan Touch Plus (Cat. No. 1640).

SPECIMEN COLLECTION AND HANDLING

Specimen: Fresh serum or urine is the specimen of choice.

Storage and Stability: If storage is necessary, samples may be stored covered at 2 to 8°C for up to 72 hours.

Interfering Factors:

- Evaporation of uncovered specimens may cause inaccurate results.
- Use of plasma will cause a fibrinogen band to appear as a distinct narrow band between the beta and gamma fractions. Although fibrinogen does not react with the antisera provided in the kit, fibrinogen may potentially adhere to the gel matrix resulting in nonspecific banding patterns.

PROCEDURE

Materials provided: The following materials needed for the procedure are contained in the SPIFE Nexus IFE-6 Kit. Individual items are not available.

SPIFE IFE-6 Gels (10)	SPIFE Nexus Pipette Wash	1 vial
SPIFE Nexus Violet (1 vial)	Fixative	1 vial
Clear Wash (1 pkg)	IgG	1 vial
Citric Acid Destain (1 pkg)	IgA	1 vial
SPIFE Nexus Blotter D (10)	IgM	1 vial
SPIFE Blotter C (10)	Kappa	1 vial
Serrated Blade Applicator Kit, 12 Sample (30)	Lambda	1 vial

Materials provided by Helena Laboratories but not contained in the kit:

Item	Cat. No.
SPIFE Nexus Analyzer	1650
QuickScan Touch Plus	1640
SPIFE IFE-3/6 Dispo Cup Tray	3377
SPIFE IFE-3/6 Dispo Sample Cups	3368

Gel Block Remover	1115
SPIFE Nexus Cassette	2580
SPIFE Nexus Applicator Templates	2570
SPIFE Nexus Applicator Blade Weights	2572
SPIFE Nexus Dispo Stain Cups	2575
Pos ID Barcode Labels for Touch & SPIFE Nexus Systems	1696
REP Prep	3100
SPIFE Nexus Reagent Roller	2583
SPIFE Nexus Ready Run Kit	2582
SPIFE Nexus Antisera Spreader Tips	2574
SPIFE Nexus Carbon Electrode Insert	2576

Materials and Supplies Needed but not Supplied:

0.85% saline

STEP-BY-STEP METHOD

I. Sample Preparation

Desired dilutions are operator programmable and may be individually set. Available dilutions are Neat; 1 in 2; 1 in 3; 1 in 4; 1 in 5; 1 in 6; 1 in 7; 1 in 8; 1 in 10; 1 in 12; 1 in 14; 1 in 16; 1 in 18, and 1 in 20 with options ranging from Neat to 1:20. See the SPIFE Nexus Operator's Manual for additional instructions. More concentrated samples are more likely to prozone while the more dilute samples may not exhibit desired sensitivity.

Serum

The SPIFE Nexus automatically samples and dilutes the specimens as follows:

SP = 1:3 (1 part serum with 2 parts 0.85% saline)
IgG = 1:5 (1 part serum with 4 parts 0.85% saline)
IgA = 1:5
IgM = 1:5
κ = 1:5
λ = 1:5

Urine

Urine samples may be run diluted or unconcentrated. However, to achieve higher sensitivity, samples may be concentrated. Shake samples to homogenize. Centrifuge desired volume at 2000 x g for 5 minutes. Remove supernatant and concentrate as follows:

Total Protein (mg/dL)	Conc. Factor
<50	100x
50-100	50x
100-300	25x
300-600	10x
>600	5x

The SPIFE Nexus automatically samples and dilutes the specimens as follows:

SP = Neat
IgG = Neat
IgA = Neat
IgM = Neat
κ = Neat
λ = Neat

For urine specimen volumes measuring less than 500 uL contact Technical Services for instructions on manual loading.

II. SPIFE Nexus Preparation

- Fill the designated bottles with 0.85% saline, deionized water, destain and Clear Wash Solution.
- Turn on the SPIFE Nexus. Click on the SPIFE Nexus icon to initialize it.
- If this is the first test of the day, prime the instrument according to the instructions in the SPIFE Nexus Operator's Manual.
- Open the main door of the instrument and prepare the items onboard the instrument.
 - Ensure that each of the following items are in their respective onboard storage locations: **Platen Cover** with the Carbon Electrode Insert, **Antisera Spreader Tip**, and **Dryer Cover** with the red sticker toward the back of the instrument.
 - Onboard Reagent Chiller**
 - Uncap and place the Antisera and Protein Fixative into the labeled onboard Reagent Chiller positions.
 - Uncap and place the Saline Diluent into Diluent Well 1, and Pipette Wash into Diluent Well 2 of the Reagent Chiller.

3. Sample Cup Tray

- Prepare the sample cup tray with the appropriate Disposable Sample Cups. Slide the sample cups into the cup tray.
- Place the cup tray onto the sample tray platform.

4. Stain/Reagent Dispenser

- Fill two Stain Cups each with 400 uL of SPIFE Nexus Violet stain. Place the Stain Cups in the outer two slots of the Stain/Reagent Dispenser. **NOTE: Do not add stain to the middle slot.**
- Place a clean Reagent Roller bar between the hooks on the Stain/Reagent Dispenser.

5. Consumables Tray

- Slide the Consumables Tray forward from its home position.
- Prepare the Applicator Holder
 - Place an IFE-6 Applicator Template on top of the Applicator Holder. Place Applicator Blades in the designated slots corresponding to the sample cups loaded within the sample tray. **NOTE: The Applicator Blades will only fit into the slots in the Applicator Holder one way; do not try to force the Applicator Blades into the slots.**
 - Place the Applicator Blade Weights on top of the Applicator Blades with the thick side facing the front of the instrument.
- Prepare the Blotter Holder
 - Flip the Blotter Holder upside down so the foam surface is upright and place the Blotter Guide around the foam to assist in blotter placement. Locate the double-sided tape on the SPIFE Nexus Blotter D and remove the adhesive backing. Adhere the blotter to the foam surface of the Blotter Holder. Remove the Blotter Guide and place the Blotter Holder back in the designated location within the Consumables Tray with the green dot facing toward the front of the instrument.
- Slide the Consumables Tray into position in the back of the instrument.

6. Gel Cassette

- Place the bottom half of the Gel Cassette on the electrophoresis platen with the two pins lined up on the left side.
- Dispense 2 mL of REP Prep on the platen.
- Remove the gel from the protective packaging and discard the overlay.
- Using a SPIFE Blotter C, gently blot the entire gel. Discard the blotter.
- Place the left edge of the gel into the bottom of the cassette fitting the round hole over the upper pin and the obround hole over the lower pin. Gently lay the gel down over the REP Prep making sure no bubbles remain under the gel.
- Place the top half of the Gel Cassette over the gel. Make sure the 2D barcode is located in the upper right corner of the cassette.
- Place a Positive ID Barcode Label on the upper right hand side of the gel backing. Select the barcode that starts with the letter "G".

E. Close the main door of the instrument.

F. Load the correct number of uncapped patient sample test tubes into test tube racks and place racks within the tube transport area.

III. Electrophoresis Parameter Setup

Click the Set Up button and select **Test Type: Serum Immunofixation (IFE) 6 (Acid Violet)** or **Urine Immunofixation (IFE) 6 (Acid Violet)** from the Test tab and click the Edit button. Check the programmed parameters for each of the following processes. Use the Cancel buttons to exit the Set Up menus if no changes are made.

A. Serum

Sample Loading

Sample Load Enabled:	Yes
Tray:	IFE 6 Disposable 3377-81541041
Volumes:	
Row 1	
Required Volume (uL):	19
Dispense Volume (uL):	40
Row 2	
Required Volume (uL):	19
Dispense Volume (uL):	40
Row 3	
Required Volume (uL):	19
Dispense Volume (uL):	40
Primary Wash Time (sec):	1
Primary Wash Cycles:	0
Secondary Wash Time (sec):	1
Secondary Wash Cycles:	0
Samples per Gel:	6
Lanes per Sample:	6
Dilution Mix Cycles:	0
Cover gel while loading samples:	Yes
Dilutions:	
Sample:	SP G A M K L
	1:3 1:5 1:5 1:5 1:5 1:5

Gel Preparation

Stain Type:	Acid Violet
Applicator Load 1	Applicator Load Enabled: Yes
	Applicator Load Time (mm:ss): 00:02
	Applicator Load Speed (mm / sec): 18
	Applicator Load Offset (mm from center): -2
Applicator Load 2	Applicator Load Enabled: Yes
	Applicator Load Time (mm:ss): 00:02
	Applicator Load Speed (mm / sec): 18
	Applicator Load Offset (mm from center): -2

Applicator Load 3	Applicator Load Enabled:	Yes
	Applicator Load Time (mm:ss):	00:02
	Applicator Load Speed (mm / sec):	18
	Applicator Load Offset (mm from center):	-2
Applicator Load 4	Applicator Load Enabled:	Yes
	Applicator Load Time (mm:ss):	00:15
	Applicator Load Speed (mm / sec):	18
	Applicator Load Offset (mm from center):	-2
Sample Application	Sample Application Enabled:	Yes
	Apply Time (mm:ss):	00:30
	Absorption Time (mm:ss):	00:00
	Application Offset (mm from center):	0

Proceed to Step III. C.

B. Urine

Sample Loading

Sample Load Enabled:	Yes
Tray:	IFE 6 Disposable 3377-81541041
Volumes:	
Row 1	
Required Volume (µL):	19
Dispense Volume (µL):	40
Row 2	
Required Volume (µL):	19
Dispense Volume (µL):	40
Row 3	
Required Volume (µL):	19
Dispense Volume (µL):	40
Primary Wash Time (sec):	1
Primary Wash Cycles:	0
Secondary Wash Time (sec):	1
Secondary Wash Cycles:	0
Samples per Gel:	6
Lanes per Sample:	6
Dilution Mix Cycles:	0
Cover gel while loading samples:	Yes
Dilutions:	
Sample:	SP G A M K L
	Neat Neat Neat Neat Neat Neat

Gel Preparation

Stain Type:	Acid Violet
Applicator Load 1	Applicator Load Enabled: Yes
	Applicator Load Time (mm:ss): 00:02
	Applicator Load Speed (mm / sec): 18
	Applicator Load Offset (mm from center): -2
Sample Application 1	Sample Application Enabled: Yes
	Apply Time (mm:ss): 00:30
	Absorption Time (mm:ss): 00:00
	Application Offset (mm from center): 0
Applicator Load 2	Applicator Load Enabled: Yes
	Applicator Load Time (mm:ss): 00:02
	Applicator Load Speed (mm / sec): 18
	Applicator Load Offset (mm from center): -2
Sample Application 2	Sample Application Enabled: Yes
	Apply Time (mm:ss): 00:30
	Absorption Time (mm:ss): 00:00
	Application Offset (mm from center): 0
Applicator Load 3	Applicator Load Enabled: Yes
	Applicator Load Time (mm:ss): 00:02
	Applicator Load Speed (mm / sec): 18
	Applicator Load Offset (mm from center): -2
Sample Application 3	Sample Application Enabled: Yes
	Apply Time (mm:ss): 00:30
	Absorption Time (mm:ss): 00:00
	Application Offset (mm from center): 0

C. Serum and Urine

Electrophoresis	Electrophoresis Enabled:	Yes
	Operating Mode:	Constant Voltage
	Voltage (volts):	650 V
	Temperature (°C):	20°C
	Time (hh:mm:ss):	00:06:30
Fixative/Antiserum Application	Fixative / Antiserum Application Enabled:	Yes
	Absorption Time (mm:ss):	00:01
	Lane 1 Position (mm from calibration point):	9.60
	Row 1 Start (mm from calibration point):	117.00
	Row 1 End (mm from calibration point):	90.00

	Lane to Lane Spacing (mm):	10.32
	Row to Row Spacing (mm):	35.60
Pre-Dry / Incubate 1	Incubate / Pre-Dry Enabled:	Yes
	Temperature (°C):	21°C
	Time (hh:mm:ss):	00:02:00
Blot	Blot Enabled:	Yes
	Temperature (°C):	50°C
	Time (hh:mm:ss):	00:04:00
Pre-Dry / Incubate 2	Incubate / Pre-Dry Enabled:	Yes
	Temperature (°C):	50°C
	Time (hh:mm:ss):	00:06:00
Buffer Wash	Process Type:	Buffer Wash
	Buffer Wash Enabled:	Yes
	Recirculate:	Recirculate Top to Bottom
	Recirculate Speed:	10
	Time (hh:mm:ss):	00:05:00
	Pulse Cycles:	5
	Pulse Time (sec):	1
Reagent / Stain Application	Reagent / Stain Application Enabled:	Yes
	Spread Cycles:	3
	Half Cycle Spread:	Yes
	Temperature (°C):	37°C
	Absorption Time (mm:ss):	00:01
	Pour Position (mm from cal. point):	0.00
	Spread Start (mm from cal. point):	0.00
	Spread Length (mm):	124.46
	End of Roll Pause (sec):	0
Destain 1	Process Type:	Destain
	Destain Enabled:	Yes
	Recirculate:	Recirculate Top to Bottom
	Recirculate Speed:	10
	Time (hh:mm:ss):	00:01:00
	Pulse Cycles:	3
	Pulse Time (sec):	1
Dry 1	Dry Enabled:	Yes
	Temperature (°C):	65°C
	Time (hh:mm:ss):	00:02:00
Destain 2	Process Type:	Destain
	Destain Enabled:	Yes
	Recirculate:	Recirculate Top to Bottom
	Recirculate Speed:	10
	Time (hh:mm:ss):	00:01:00
	Pulse Cycles:	3
	Pulse Time (sec):	1
Dry 2	Dry Enabled:	Yes
	Temperature (°C):	65°C
	Time (hh:mm:ss):	00:02:00
Scan	Scan Enabled:	Yes

Instrument Setup

Cassette:	C01: Standard Cassette
Applicator Template:	A01: IFE 3, 6, 9 / CK / SPE / Split Beta (60)
Reagent Applique:	R01: Antisera Storage
Electrophoresis Plate:	E01: Standard 2 Carbon Electrodes
Blotter Holder:	B01: Standard Blotter

IV. Automated Gel Electrophoresis

1. Click the Start button on the menu bar. Select the appropriate **test name** from the drop down menu. Ensure the toggles for all Run Processes are set to "Yes" and click the Start Run button. The analyzer will load samples when appropriate, apply samples, electrophoresis, immunofix, wash, stain, destain, dry and scan the gel.
2. After scanning, the Gel Cassette with the finished gel will be located in the scanner port of the front side of the instrument. If gel storage is required, remove and discard the two gel blocks.
3. After every test: discard the used blotters, Applicator Blades, Stain Cups and sample cups as biohazardous waste. Clean any residual stain from the electrophoresis platen, Gel Cassette and the Reagent Roller bar. For daily, weekly, and monthly maintenance reference the SPIFE Nexus Operator's Manual.

Qualitative Evaluation: The SPIFE IFE-6 Gel will be automatically scanned. Refer to the QuickScan Touch Plus Operator's Manual for scanning parameters.

Stability of End Product: The completed, stained and dried immunofixation gel is stable for an indefinite period of time.

INTERPRETATION OF RESULTS

Normal sample: No monoclonal present, immunofixation lanes have faint diffused stain or blush of color or variable intensity that reflects the normal distribution of immunoglobins.

Polyclonal: No monoclonal present, a diffuse increase in at least one heavy chain and both light chains.

Monoclonal: A monoclonal protein is characterized by a well-defined restricted band in a heavy chain lane with a corresponding band in a light chain lane. The monoclonal protein band on the immunofixation pattern will occupy the same migration position and shape as the monoclonal band on the reference protein electrophoresis pattern. The abnormal protein is identified by the corresponding antiserum used. Because of the increased sensitivity of the procedure, it is not uncommon to see a fixed band that is not visible in the serum protein procedure. The majority of monoclonal proteins migrate in the cathodic (gamma) region of the protein pattern. However, due to their abnormality, they may migrate anywhere within the protein electrophoresis pattern.

When low concentrations of M-protein are present, the immunofixation band may appear on the stained background of the polyclonal immunoglobulin. A stained background may also appear when the M-protein is present along with a large polyclonal increase.

Other:

Reaction with light chain antisera only could indicate either a free light chain gammopathy or (rarely) IgD or IgE gammopathy. (See Further Testing)

Reaction with heavy chain antisera may indicate (rarely) heavy chain disease or an atypical light chain.⁷

Multiple M-Proteins - On rare occasion biconal (two M-proteins) or oligoclonal (more than two) patterns may occur.

Fc fragments - Proteolysis of the immunoglobulin into Fac and Fc fragments may occur either in vivo or in vitro. The Fc fragments may maintain antigen antibody specificity to a heavy chain only, usually migrate in the alpha-2 area, and appear monoclonal in nature, with no corresponding band in the light chain lane. Fab fragments react with heavy and light chain antisera and are polyclonal in appearance.³

Restricted heterogeneity/Ladder light chains⁹ - A pattern of multiple regularly spaced restricted bands reacting with light chain antisera may be seen in urine with a high concentration of polyclonal light chains.⁴

Cathodal bands - An extremely cathodal band present in the SP lane that does not react with antisera is consistent with lysozyme. Elevated lysozyme may be seen in monocytic and monomyelocyte leukemias.⁵

For an in-depth discussion of IFE interpretation, call Helena Laboratories toll free and request the free publication "An Immunofixation Tutorial" Book F.

Further Testing Required

Specimens containing a band on serum protein electrophoresis suggestive of a monoclonal protein, but which do not react with IgG, IgA or IgM antisera, may require further testing as follows:

1. Serum samples which have a precipitin band with Kappa or Lambda Light Chain Antisera but none corresponding with IgG, IgA or IgM antisera may have a free light chain or they may have an IgD or IgE monoclonal protein. Such sera should be tested with Immunofix IgD and IgE antisera.
2. A CRP band may be detected in patients with acute inflammatory response.⁶ CRP appears as a narrow band on the most cathodic end of the high resolution agarose protein electrophoresis pattern. Evaluated Alpha-1 Antitrypsin and haptoglobin (acute phase proteins) are supportive evidence for the presence of a CRP band. Patients with a CRP band will have a positive CRP by latex agglutination or an elevated quantitative CRP.

LIMITATIONS

1. Antigen excess will occur if there is not a slight antibody excess or antigen/antibody equivalency at the site of precipitation. Antigen excess in IFE is usually due to a very high level of immunoglobulin in the patient sample. The dissolution of immunoprecipitation is manifested by a loss of protein at the point of highest antigen concentration, resulting in staining at the margins of the band, while leaving the central area with little demonstrable protein stain. In this case, it may be necessary to adjust the protein content of the sample by dilution. Electrophoresing excessive amounts of antigen decreases resolution and requires higher concentrations of antibody. For optimum separation and sufficient intensity for visual detection, care must be taken in adjusting antibody content, sample concentration, time and voltage.
2. Monoclonal proteins may occasionally adhere to the gel matrix, especially cryoglobulins or IgM. These bands will appear in all five antisera reaction areas of the gel. However, where the band reacts with the specific antisera for its heavy chain and light chain, there will be a marked increase in size and staining activity, allowing the band to be identified.
3. An application artifact may appear as a fine clear line (negative space) that may be visible to a faint degree across the entire gel in the beta region. This can on occasion cause the edge of a normal blush to appear slightly blunted.
4. Therapeutic monoclonal antibodies may be used in the treatment of multiple myeloma as well as various other malignancies or medical conditions. If present in sufficient concentration, a humanized therapeutic monoclonal antibody will react with antisera in a manner comparable to a pathologic monoclonal protein.^{7,8}

5. Light chains associated with IgA or IgD heavy chains may on rare occasion be difficult to visualize. The structure of some IgA dimers may potentially block the light chain epitopes, decreasing the antisera reaction. This occurs more commonly with lambda than kappa.^{9,10}
6. It is possible that not all monoclonals are detected by immunofixation. Not all clinically significant monoclonal gammopathies will display a distinct band detectable by serum protein electrophoresis.¹¹

PERFORMANCE CHARACTERISTICS

Serum and urine samples containing IgG, IgA, IgM, Kappa light chain and Lambda light chain monoclonal proteins were tested using the SPIFE Touch and SPIFE Nexus instruments. The test results showed complete concordance between instruments.

BIBLIOGRAPHY

1. Cawley, L.P., Minard, B.J., Tourtellotte, W.W., Ma, B.I., & Chelle, C. Immunofixation electrophoretic techniques applied to identification of proteins in serum and cerebrospinal fluid. *Clin Chem.* 1976; 22(8), 1262-1268.
2. Ritchie RF, Smith R. Immunofixation. III. Application to the study of monoclonal proteins. *Clin Chem.* 1976; 22(12):1982-1985.
3. Keren DF. In: *Protein Electrophoresis in Clinical Diagnosis*. American Society for Clinical Pathology Press; 2012:125-134.
4. Macnamara EM, et al. *Clin Chem.* 1991; 37(9): 1570-1574.
5. Ritzmann SE, editor. *Protein Abnormalities Volume 3. Proteins in Bodily Fluids, Amino Acids, and Tumor Markers: Diagnostic and Clinical Aspects*. New York (NY): Allen R Liss; 1983; 130.
6. Jeppson JO, Laurell CB, Franzén B. Agarose gel electrophoresis. *Clin Chem.* 1979; 25(4):629-638.
7. Keren DF. Therapeutic complications: A caveat for M-protein detection. *The Journal of Applied Laboratory Medicine.* 2016; 1(4):342-345.
8. McCudden CR, Jacobs JFM, Keren D, Caillon H, Dejoie T, Andersen K. Recognition and management of common, rare, and novel serum protein electrophoresis and immunofixation interferences. *Clin Biochem.* 2018; 51:72-79.
9. Yu M, Bruns DE, Katzmann JA, Silverman LM, Murray DL. Restricted IGG-kappa and free alpha-heavy-chain bands in an asymptomatic 62-year-old man. *Clin Chem.* 2018; 64(2):265-268.
10. Hashimoto N, Chandor S, Mandy W, Yokoyama M. Atypical IgA with hidden light chain. *Clin Exp Immunol.* 1970; 6(6):941-949.
11. Kyle RA, Gertz MA, Witzig TE, et al. Review of 1027 patients with newly diagnosed multiple myeloma. *Mayo Clinic Proceedings.* 2003; 78(1):21-33.

For Sales, Technical and Order Information and Service Assistance, call 800-231-5663 toll free.

Helena Laboratories warrants its products to meet our published specifications and to be free from defects in materials and workmanship. Helena's liability under this contract or otherwise shall be limited to replacement or refund of any amount not to exceed the purchase price attributable to the goods as to which such claim is made. These alternatives shall be buyer's exclusive remedies. In no case will Helena Laboratories be liable for consequential damages even if Helena has been advised as to the possibility of such damages.

The foregoing warranties are in lieu of all warranties expressed or implied including, but not limited to, the implied warranties of merchantability and fitness for a particular purpose.



Beaumont, Texas USA 77707

Pro. 78
4/22(2)